


# Technical Data Sheet

# Rosenberger

RPC-3.50

Calibration Kit  
Jack

03K30R-MSOS3



All dimensions are in mm; tolerances according to ISO 2768 m-H

## Interface

According to  
Mechanically compatible with

IEC 61169-23  
RPC-2.92 and SMA

## Contents and Documentation

This kit is delivered with

- **Standard Definitions Card**  
Printed Standard Definitions that can be used on nearly all Vector Network Analyzers
- **Test Results Documentation**
- **Hard Shell Case**
- **Protection Caps**

## Material and plating

### Connector parts

Center conductor  
Outer conductor  
Body  
Dielectric  
Substrate

### Material

CuBe  
Stainless steel  
Aluminum  
PS  
Al<sub>2</sub>O<sub>3</sub>

### Plating

Gold, min. 1.27 µm, over nickel  
Passivated  
black anodized

**Electrical data**

Frequency range DC to 26.5 GHz

**Open**

Error from nominal phase<sup>1</sup>  
 $\leq 1.0^\circ$ , DC to 4 GHz  
 $\leq 2.0^\circ$ , 4 GHz to 8 GHz  
 $\leq 3.0^\circ$ , 8 GHz to 26.5 GHz

**Short**

Error from nominal phase<sup>2</sup>  
 $\leq 1.0^\circ$ , DC to 4 GHz  
 $\leq 2.0^\circ$ , 4 GHz to 8 GHz  
 $\leq 3.0^\circ$ , 8 GHz to 26.5 GHz

**Load**

Return loss  
 $\geq 40.0$  dB, DC to 4 GHz  
 $\geq 35.0$  dB, 4 GHz to 8 GHz  
 $\geq 30.0$  dB, 8 GHz to 26.5 GHz

DC Resistance 50  $\Omega \pm 0.5 \Omega$   
 Power handling (at 25 °C, sea level)  $\leq 0.5$  W, derate by 0.005 W/K

<sup>1</sup> The nominal phase is defined by the Offset Delay, the Offset Loss and the Fringing Capacitances

<sup>2</sup> The nominal phase is defined by the Offset Delay, the Offset Loss and the Short Inductance

**Mechanical data**

Mating cycles  $\geq 500$   
 Maximum torque 1.70 Nm  
 Recommended torque 0.90 Nm  
 Gauge 0.00 mm to 0.08 mm

**General standard definitions**

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

**Open**

Offset  $Z_o$  / Impedance /  $Z_o$  50  $\Omega$   
 Offset Delay 33.356 ps  
 Length (electrical) / Offset Length 10.00 mm  
 Offset Loss 2.20 G $\Omega$ /s  
 Loss 0.0127 dB/ $\sqrt{\text{GHz}}$   
 Fringing Capacitances  
 $C_0 = -17.500 \times 10^{-15}$  F / -17.500 fF  
 $C_1 = -2000.0 \times 10^{-27}$  F/Hz / -2.0000 fF /GHz  
 $C_2 = 140.00 \times 10^{-36}$  F/Hz / 0.1400 fF /GHz<sup>2</sup>  
 $C_3 = -2.7000 \times 10^{-45}$  F/Hz / -0.0027 fF /GHz<sup>3</sup>

# Technical Data Sheet

# Rosenberger

RPC-3.50

Calibration Kit  
Jack

03K30R-MSOS3

## Short

Offset $Z_o$ / Impedance / $Z_o$	50 $\Omega$
Offset Delay	33.356 ps
Length (electrical) / Offset Length	10.00 mm
Offset Loss	2.36 G $\Omega$ /s
Loss	0.0137 dB/ $\sqrt{\text{GHz}}$
Short Inductance	$L_0 = -44.000 \times 10^{-12} \text{ H} \quad / \quad -44.000 \text{ pH}$
	$L_1 = 3700.0 \times 10^{-24} \text{ H/Hz} \quad / \quad 3.7000 \text{ pH /GHz}$
	$L_2 = -250.00 \times 10^{-33} \text{ H/Hz}^2 \quad / \quad -0.2500 \text{ pH /GHz}^2$
	$L_3 = 5.0000 \times 10^{-42} \text{ H/Hz}^3 \quad / \quad 0.0050 \text{ pH /GHz}^3$

## Load

Offset $Z_o$ / Impedance / $Z_o$	50 $\Omega$
Offset Delay	0.0000 ps
Length (electrical) / Offset Length	0.000 mm
Offset Loss	0.00 G $\Omega$ /s
Loss	0.0000 dB/ $\sqrt{\text{GHz}}$

## Environmental data

Operating temperature range <sup>3</sup>	+20 °C to +26 °C
Rated temperature range of use <sup>4</sup>	0 °C to +50 °C
Storage temperature range	-40 °C to +85 °C
RoHS	compliant

<sup>3</sup> Temperature range over which these specifications are valid.

<sup>4</sup> This range is underneath and above the operating temperature range, within the calibration kit is fully functional and could be used without damage

## Declaration of documentation

Standard delivery for this kit includes Test Results. The documentation issued reports which quantities were tested individually, traceable to national / international standards. Model based standard definitions of the calibration standards are reported in Agilent / Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

## Inspection interval

Recommendation	12 months
----------------	-----------

## Packing

Standard	1 pce in bag
Weight	24.7 g/pce

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
Marcel Panicke	07.01.16	Markus Müller	29.04.20	e00	19-2083	Marion Striegler	29.04.20

Rosenberger Hochfrequenztechnik GmbH & Co. KG  
P.O.Box 1260 D-84526 Tittmoning Germany  
www.rosenberger.de

Tel. : +49 8684 18-0  
Email : info@rosenberger.de

Page  
3 / 3