


IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Switch Mode Power Supply
Name and address of the applicant	VOX POWER LTD Unit 2 Redcow Interchange Estate Dublin, 22 Ireland
Name and address of the manufacturer	VOX POWER LTD Unit 2 Redcow Interchange Estate Dublin, 22 Ireland
Name and address of the factory <i>Note: When more than one factory, please report on page 2</i>	EOS POWER INDIA PVT LTD UNIT 57 SDF-II SEEPZ ANDHERI (E) MUMBAI MH 400096 INDIA <input checked="" type="checkbox"/> Additional Information on page 2
Ratings and principal characteristics	See Page 2 Vox-Power 
Trademark (if any)	
Type of Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	NEVO+600ML-abcd-zzzz, NEVO-600M-abcd-zzzz, NEVO+600M-abcd-zzzz See Page 2
Additional information (if necessary may also be reported on page 2)	<input checked="" type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1
As shown in the Test Report Ref. No. which forms part of this Certificate	E316486-D1 issued on 2018-04-19

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2018-05-09
Original Issue Date: 2016-02-23

Signature:

Jan-Erik Storgaard

Model Details:

NEVO+600ML-abcd-zzzz,NEVO-600M-abcd-zzzz,NEVO+600M-abcd-zzzz where :

a is slot A and can be 0, 1, 2, 3, 4, 5, 6, 7 or 8

b is slot B and can be 0, 1, 2, 3, 4, 5, 6, 7 or 8

c is slot C and can be 0, 1, 2, 3, 4, 5, 6, 7 or 8

d is slot D and can be 0, 1, 2, 3, 4, 5, 6, 7 or 8

0=No module (blanking plate fitted)

1=OP1 output module

2=OP2 output module

3=OP3 output module

4=OP4 output module

5=OP5 output module

6=OP6 output module

7=OP7 output module

8=OP8 output module

May also be followed by up to 30 alphanumeric characters zzzz indicating non-safety related information. V1, V2, V3, V4 indicates non-safety related information.

Factories:

PANYU TRIO MICROTRONICS CO LTD

SHIJI INDUSTRIAL ESTATE DONGYONG NANSHA GUANGZHOU GUANGDONG 511453

CHINA

Ratings:

Input:

NEVO-600M;

100-240Vac, 4.5 A, 50-60 Hz

120-300 Vdc, 4.5 A

NEVO+600M and model NEVO+600ML;

100-240 Vac, 6 A, 50-60 Hz

120-300 Vdc, 6 A

Output:

Max. total output power 600 W (model NEVO+600ML: 450W)

Output modules (max. ratings)

OP1; 1.5 - 7.5 Vdc, 25 A, 125 W

OP2; 4.5 - 15 Vdc, 15 A, 150 W

OP3; 9- 30 Vdc, 7.5 A, 150 W

OP4; 18-58 Vdc, 3.75 A, 150 W

OP5; 3.3 - 15 Vdc, 5 A, 75 W and/or

3.3 - 15 Vdc, 5 A, 75 W

OP6; 1.8 - 5 Vdc, 5 A, 25 W and/or

1.8 - 5 Vdc, 5 A, 25 W

OP7; 5 - 15 Vdc, 5 A, 75 W and/or

1.8 - 5 Vdc, 5 A, 25 W

OP8= dual output 2x 24Vdc, 3.125A (2 x 75W) 150W total

(OP5, OP6 and OP7 may have either 1 or 2 outputs)

All output module current and power ratings are the maximum allowable.

Additional Information:

Risk Management was not included in this investigation.

Additionally evaluated to EN 60601-1:2006/ A1:2013/ A12:2014.

National Differences specified in the CB Test Report.

The original report was modified to include the following changes/additions:

Correct output module OP5 output rating.

Additional information (if necessary)

UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/nbnames

Date: 2018-05-09

Original Issue Date: 2016-02-23

Signature:

Jan-Erik Storgaard