

Dear: Distribution Partner

TDK Corporation
Electronic Components Business Company
Piezo & Protection Devices BG

Subject: Nebulizer product - Information of change of a screw lock

To whom it may concern:

We inform that we change a screw lock of nebulizer unit.
Please confirm the contents of change and the schedule.

<Reason of change>

Adhesive of the screw lock used in our nebulizer unit is subject to the following VOC standards.

Overview of VOC Standards:

As for chemicals including adhesives, coatings, inks and cleaning agent which are produced and sold in China and/or import to China from overseas, that exceed VOC content limits stipulated in GB standards (Mandatory standards) is prohibited to produce and sale in China and/or import to China from overseas.

(Issued in March 4th, 2020 / Posted in SAC Website in March 6th, 2020)

*VOC: Volatile Organic Compounds

Therefore, we will change adhesive for screw lock in order to comply to VOC standards.

<Contents of change>

We will change to a screw lock which comply with VOC standards.

While there is a change in appearance color, the fixed strength is equivalent.

There is no a change in the shape, characteristics and specs of the nebulizer unit associated with this change.

Please refer to the attachment (P2) "1. Contents of change" for the details.

<Applicable product to replace>

Applicable product is a nebulizer unit.

Please refer to the attachment (P3) "2. Applicable product list" for the details.

<Schedule for change>

We plan to replace sequentially from April 2021.

As for the change of adhesive, since it doesn't affect to the characteristics and specs of the product, We would like to make a running change.


We appreciate your understanding and cooperation.

Sincerely yours,
TDK

1. Contents of change

1-1 : Appearance

We will change screw lock applied to the screw.
Please refer to the below pictures.



1-2 : Screw fixed strength

The screw fixing strength is the same before and after change.
Please refer to the table below.



