


# Product Change Notice

## 3M™ Ground Pro Ground Integrity Meter CTM051



Issue Number: 1413  
Date: Oct. 10, 2014

### Product Change

Effective immediately, the Ground Pro Ground Integrity Meter CTM051 will have the following changes. This impacts all pending orders and new orders.

Items	Before	After	Action
<p><b>Updated User Guide, Product Data Sheet and Product Brochure</b></p> <p>Updates have been made to provide clarity of the product intent of use and updated safety warnings</p>		<p>See attached:</p> <p>User Guide Product Data Sheet Product Bulletin</p>	<p>An updated User Guide, Product Data Sheet and Product Bulletin are available on the website and are also attached within this e-mail notification.</p> <p>Please ensure an updated User Guide is provided to your customers that have previously purchased a Ground Pro Integrity Meter. <a href="http://3MStatic.com">3MStatic.com</a></p>
<p><b>New Product Label:</b></p> <p>Updated important safety statements</p>	See attached picture of existing product label	See attached picture of new label	<p>Notify your customers that have previously purchased a Ground Pro Integrity Meter that the product safety labeling has changed and that New Labels are available from 3M.</p> <p>Request new labels from 3M by contacting your 3M Sales Representative or sending an e-mail to: <a href="mailto:3MStaticOrders@mmm.com">3MStaticOrders@mmm.com</a></p>
<p><b>New Product Case:</b></p> <p>All new shipments effective immediately will have an updated product case.</p>	<p>1. UL Flame rating: HB 2. Battery door without Screw</p> <p>See attached picture of existing battery door latch (without screw)</p>	<p>1. UL Flame rating: V0 2. Battery door with Screw</p> <p>See attached picture of new battery door latch (with screw)</p>	<p>Note to existing and new customers:</p> <ul style="list-style-type: none"> <li>Do not replace battery while device is in use</li> <li>Do not operate if the battery door is open</li> <li>Do not use on 3-phase equipment</li> </ul>

# Product Change Notice

## 3M™ Ground Pro Ground Integrity Meter CTM051

### Existing Product Label



### New Product Label



### Existing Battery Door Latch (no Screw)



### New Battery Door Latch (with Screw)



# Product Change Notice

## 3M™ Ground Pro Ground Integrity Meter CTM051

### Contact Information

Seokhun Kim, Static Instrument Marketing Manager  
82 2 3771 4914  
skim5@mmm.com

3M is a trademark of 3M Company.

### Important Notice

All statements, technical information, and recommendations related to 3M's products are based on information believed to be reliable, but the accuracy or completeness is not guaranteed. Before using this product, you must evaluate it and determine if it is suitable for your intended application. You assume all risks and liability associated with such use. Any statements related to the product which are not contained in 3M's current publications, or any contrary statements contained on your purchase order shall have no force or effect unless expressly agreed upon, in writing, by an authorized officer of 3M.

### Warranty; Limited Remedy; Limited Liability.

This product will be free from defects in material and manufacture for a period of one year from the time of purchase. **3M MAKES NO OTHER WARRANTIES INCLUDING, BUT NOT LIMITED TO, ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.** If this product is defective within the warranty period stated above, your exclusive remedy shall be, at 3M's option, to replace or repair the 3M product or refund the purchase price of the 3M product. **Except where prohibited by law, 3M will not be liable for any indirect, special, incidental or consequential loss or damage arising from this 3M product, regardless of the legal theory asserted.**



Electronics Materials Solutions Division  
Static Control Products

926 JR Industrial Drive  
Sanford, NC 27332-9733  
1-866-722-3736  
www.3Mstatic.com

© 3M 2014 All Rights Reserved.